



SUCCEEDER®

# SF-400

## Semi Automated Coagulation Analyzer




ISO

ISO13485:2016



No.Q5 076148 0008 Rev.03



# SF-400

Semi Automated Coagulation Analyzer

## Specification

<b>Assay</b>	Viscosity-based (mechanical) clotting assay
<b>Testing Item</b>	PT, APTT, FIB, TT, Protein C, Protein S, HEP, LMWH, Factor II / V / VII / VIII / IX / X / XI / XII.
<b>Incubation Channel</b>	16
<b>Reagent Position</b>	4
<b>Stirring Position</b>	1
<b>Display</b>	LCD with adjustable brightness
<b>Printer</b>	Internal thermal printer.
<b>Port</b>	RS232, LIS/HIS supported.
<b>Transmission</b>	HIS/LIS supported
<b>Testing Operation</b>	Automatically or manually, pipette compatible.
<b>Incubation Time</b>	0-999s
<b>Incubation Timer</b>	4 individual timers, with counting down display and alarm.
<b>Power Supply</b>	AC100V~250V, 50/60HZ
<b>Analyzer Dimension</b>	410*390*130 (L*W* H, mm)
<b>Analyzer Weight</b>	15 kg

仪器 · 试剂 · 服务

**Thrombosis  
Hemostasis**



### BEIJING SUCCEEDER TECHNOLOGY INC.

Addr: Block A, Building 1, Building 2, No.19 Baili Street, Daxing District, Beijing, China. 102600

Tel: +86-10-8928 1758

E-mail: sales@succeeder.com.cn

Fax: +86-10-8928 1700

Website: www.succeeder.com